

CN01 - Clobazam, Norclobazam

Analyte	Analyst	Method	Result	Units	Z score (** z' score)	Assigned Value	Ux AV	SDPA	Exp.SDPA	No of results	Median	Mean	Robust SD	SD
CN1A - Clobazam	Lab Result	LCMS & MS/MS	127.00	µg/L	-0.58	136.70	2.03	16.691	N/A	19	136.35	139.60	11.419	12.948
CN1B - Clobazam	Lab Result	LCMS & MS/MS	676.00	µg/L	0.04	672.87	9.01	72.857	N/A	19	672.87	678.57	34.005	67.890
CN1A - Norclobazam	Lab Result	LCMS & MS/MS	205.00	µg/L	-0.74	232.00	5.06	36.617	N/A	14	232.00	254.54	23.135	60.759
CN1B - Norclobazam	Lab Result	LCMS & MS/MS	4,005.00	µg/L	-0.79	4,400.28	54.10	497.535	N/A	14	4,415.18	4,362.65	228.115	261.621

PS23 - Zuclopenthixol

Analyte	Analyst	Method	Result	Units	Z score (** z' score)	Assigned Value	Ux AV	SDPA	Exp.SDPA	No of results	Median	Mean	Robust SD	SD
Zuclopenthixol	Lab Result	LCMS & MS/MS	33.50	µg/L	-1.17	41.37	0.95	6.719	N/A	31	41.38	42.24	4.034	6.293

** Please note, participant performance for this analyte has been assessed using a z' score, rather than a z score, in order to account for the measurement uncertainty of the assigned value which is not negligible when compared to the SDPA.