

CN01 - Clobazam, Norclobazam

Analyte	Analyst	Method	Result	Units	Z score (** Z' score)	Assigned Value	Ux AV	SDPA	Exp-SDPA	No of results	Median	Mean	Robust SD	SD
CN1A - Clobazam	Lab Result	LCMS & MS/MS	177.00	µg/L	0.95	159.42	5.13	16.267	N/A	18	158.43	160.81	29.957	35.315
CN1B - Clobazam	Lab Result	LCMS & MS/MS	720.00	µg/L	0.73	762.08	18.88	85.153	N/A	18	702.54	776.54	108.103	135.836
CN1A - Norclobazam	Lab Result	LCMS & MS/MS	189.00	µg/L	-0.03	190.03	7.88	34.638	N/A	15	188.50	186.52	23.054	33.278
CN1B - Norclobazam	Lab Result	LCMS & MS/MS	3,783.00	µg/L	0.15	3,706.87	83.09	384.222	N/A	14	3,686.83	3,652.91	363.093	594.633

PS13 - Thioridazine

Analyte	Analyst	Method	Result	Units	Z score (** Z' score)	Assigned Value	Ux AV	SDPA	Exp-SDPA	No of results	Median	Mean	Robust SD	SD
Thioridazine	Lab Result	LCMS & MS/MS	505.00	µg/L	0.29	537.25	17.20	112.744	N/A	7	534.50	530.26	25.863	34.242

PS16 - Perphenazine

Analyte	Analyst	Method	Result	Units	Z score (** Z' score)	Assigned Value	Ux AV	SDPA	Exp-SDPA	No of results	Median	Mean	Robust SD	SD
Perphenazine	Lab Result	LCMS & MS/MS	0.24	µg/L	0.14	0.28	0.02	0.200	N/A	20	0.28	0.32	0.050	0.152

** Please note, participant performance for this analyte has been assessed using a Z' score, rather than a Z score, in order to account for the measurement uncertainty of the assigned value which is not negligible when compared to the SDPA.