

OVERVIEW Z-SCORES

 UV:Satisfactory (-2 ≤ Z-score ≤ 2)
 UV: Need attention (-3 < Z-score < -2 or 2 < Z-score < 3)
 UV:Unsatisfactory (Z-score ≤ -3 or ≥ 3)
 BV: Unacceptable

Module	Parameter	Sample	Test System	Total	Assay		Method	Bivariate Analysis
Coagulation Factor - I	Factor VIII (clotting activity)	24.160	1	-0.24	-0.24		0.63	
		24.161	1	-0.12	-0.13		-0.39	
Coagulation Factor - I	Factor IX (clotting)	24.160	1	0.59	0.58		1.11	
		24.161	1	0.51	0.50		0.81	
Coagulation Factor - I	Factor XI	24.160	1	-0.15	-0.15		0.62	
		24.161	1	-0.22	-0.23		0.25	
Coagulation Factor - I	Factor XII	24.160	1	0.40	0.40		0.86	
		24.161	1	0.96	0.96		1.22	
Coagulation Factor - II	Factor II	24.162	1	-0.88	-0.88		0.53	
		24.163	1	-0.55	-0.55		0.48	
Coagulation Factor - II	Factor V	24.162	1	0.34	0.34		0.62	
		24.163	1	0.08	0.08		0.73	
Coagulation Factor - II	Factor VII	24.162	1	-0.89	-0.89		-0.69	
		24.163	1	-0.86	-0.86		-0.95	
Coagulation Factor - II	Factor X	24.162	1	0.05	0.05		-0.14	
		24.163	1	-0.28	-0.28		-0.29	
Von Willebrand Factor Par	VWF antigen	24.164	1	1.50	1.36		0.80	
		24.165	1	1.35	1.22		0.58	
Von Willebrand Factor Par	VWF activity	24.164	1	0.99	0.39		0.36	
		24.165	1	-1.46	-0.39		0.32	
Von Willebrand Factor Par	VWF Collagen Binding Assay	24.164	1	0.51	-0.13		-0.29	
		24.165	1	0.86	0.21		-0.27	
Von Willebrand Factor Par	VWF Factor VIII	24.164	1	-0.78	-0.87			
		24.165	1	-0.86	-0.95			
Factor XIII	Factor XIII activity (quantitative)	24.168	1	-0.53	-0.54		-0.43	
		24.169	1	-0.23	-0.29		-0.62	
Fibrinolysis - I	Plasminogen activity	24.170	1	-0.95	-0.95		-0.59	
		24.171	1	-1.25	-1.25		-0.60	
Fibrinolysis - I	Plasmin Inhibitor	24.170	1	-0.29	-0.29		-0.14	
		24.171	1	-0.23	-0.23		0.65	
Fibrinolysis - II	PAI-1 antigen	24.172	1	-0.74	-0.74		1.01	
		24.173	1	-0.48	-0.48		0.56	