


Distribution Number : 28
Date : Sep-24
Samples : Q6/28
Lab Number : 365

Free Kappa

Methods

Overall
Binding Site Freelite on BNII and ProSpec
Binding Site Optilite
Miscellaneous

n	Mean	SD	CV	</>
62	38.518	5.47	14.2%	0
10	49.973	9.16	18.3%	0
48	37.923	2.63	6.9%	0
5	29.240	1.33	4.5%	0



Overall Results

Bias 6.001
BIS 37.74
VIS 37.74
MRBIS 41.84
MRVIS 102.8

By Your Method

Bias 7.665
BIS 48.2
VIS 48.2
MRBIS 18.37
MRVIS 119.6

Your Result:

40.83


Binding Site Optilite

Free Lambda

Methods

Overall
Binding Site Freelite on BNII and ProSpec
Binding Site Optilite
Miscellaneous

n	Mean	SD	CV	</>
62	183.149	19.16	10.5%	0
10	207.593	33.47	16.1%	0
47	179.980	13.81	7.7%	0
5	254.700	122.97	48.3%	0



Overall Results

Bias 11.67
BIS 73.39
VIS 73.39
MRBIS 6.906
MRVIS 60.12

By Your Method

Bias 13.63
BIS 85.75
VIS 85.75
MRBIS 9.51
MRVIS 76.32

Your Result:

204.52


Binding Site Optilite

Free Light Chain Ratio

Methods

Overall
Binding Site Freelite on BNII and ProSpec
Binding Site Optilite
Miscellaneous

n	Mean	SD	CV	</>
60	0.213	0.05	22.0%	0
10	0.254	0.07	27.9%	0
43	0.214	0.02	11.5%	0
6	0.170	0.03	16.9%	0



Overall Results

Bias -6.44
BIS -40.5
VIS 40.52
MRBIS -19.4
MRVIS 107.6

By Your Method

Bias -6.57
BIS -41.3
VIS 41.35
MRBIS -4.59
MRVIS 34.11

Your Result:

0.1996

Binding Site Optilite